## Tool Configuration / Specifications

SYSTEM CONTROL PANEL
1. Digital Tachometer
2. Power ON/OFF
3. Vacuum ON/OFF
4. Start
5. Emergency Stop
6. Motor Purge Interlock
7. Vacuum Interlock
8. Resist Select
9. Dispense AUTO/OFF
10. Spindle Acceleration
11. Solvent (optional)
12. Nitrogen (optional)
13. Resist Dispense
14. Spread
15. Spin
16. Cycle Select
17. Spindle Speed
18. Digital Timers

FACILITIES AND INSTALLATION SPECIFICATIONS
Flexible

| Facilities | Tubing O.D. | 5110C |
|---|---|---|
| Nitrogen | SCFM at 60 PSI | 1/4" 20 |
| Vacuum Inches HG .20CFM | 1/4" | 20" |
| Power Volts | 115V | |
| Amps | 8 Amp | |
| Process Exhaust | | |
| SCFM at .5" of water | 4.0" | 80 |
| Cabinet Exhaust | | |
| SCFM at .5" of water | 4.0" | 100 |
| Shipping Weight | Approx. 250 lbs. | |



The information contained in this data sheet is, to our knowledge, accurate. We do not warrant the completeness or accuracy of the information contained herein. Any offer by you to purchase the equipment described in this data sheet shall be subject to the terms and conditions of sale.